



ISO-Accredited Calibration Service Documentation

Client Information	
Company	Lampert University
Address	
Phone	

Client Instrument	
Manufacturer	Brandtech
Model	Transf. S 20-200 ul
Serial #	23L02208
ID/User	R. Lampert

Laboratory Test Conditions	
Temperature (19-24 °C)	22
Relative Humidity (45-75%)	54
Bar. Pressure (kPa)	99.8
Water Den. (gm/ml)	0.9967
Density Correction (z-factor)	1.0033
Water Conductivity (µ S)	1.19
Evaporation Rate (mg)	0.002

Laboratory Test Equipment	
Test Balance Serial #	15-0043001322
Test Balance Model	#15-MCA10.6S
Test Balance Readability (mg)	0.001
Test Balance NIST Cert. Date	27/Jul/2023
Test Balance NIST Cert Due Date	31/Jul/2024

<input type="checkbox"/> O-Ring	<input type="checkbox"/> Battery
<input type="checkbox"/> Seal	<input type="checkbox"/> Shaft/Nozzle Filter(s)
<input type="checkbox"/> Friction Ring	<input type="checkbox"/> Tip Ejector
<input type="checkbox"/> Shaft/Nozzle	<input type="checkbox"/> Multi-Channel Tipcone
<input type="checkbox"/> Plunger button	<input type="checkbox"/> Housing Screw
<input type="checkbox"/> Plunger Button Cap	<input type="checkbox"/> Other (see Comments)
<input type="checkbox"/> Calibration mechanism was adjusted	

Quality Control Authorization	
QC Reviewer	MA
QC Date	05/Feb/2024
Signature	
Mike Anema, Director of Operations	

The calibration results published in this certificate were obtained gravimetrically using Grade 3 purified water, equipment manufacturer or validated substitute tips, and ISO8655 compliant test equipment that are traceable to NIST and through NIST to the International System of Units (SI). Pipettes.com certifies that the above measuring device meets or exceeds all measurement tolerances, unless otherwise noted.

Pipettes.com prohibits the reproduction of this document, except in its entirety.

Certificate ID	
02052024-23L02208	
Comments	
Calibrated new instrument.	

Service Information	
Service Date	05/Feb/2024
Certificate ID #	02052024-23L02208
Test Technician	CD
Next Due Date	31/Aug/2024

As Found Data

No 'As-Found' data taken.
Please see comments section.

Test Volumes (µL)	As Left Data		
	20	100	200
1	20.06	99.15	199.03
2	20.05	99.14	199.22
3	20.04	99.00	199.25
4	20.11	99.13	199.11
5	20.00	99.29	199.13
6	20.07	99.03	199.06
7	20.04	99.06	199.01
8	20.09	98.94	198.94
9	20.10	98.95	199.58
10	19.98	99.04	199.13
Mean (mg)	20.05	99.07	199.15
Density Corr.	1.0033	1.0033	1.0033
Mean (ul)	20.12	99.40	199.81
Accuracy (% Dev)	0.60	-0.60	-0.10
Inaccuracy Tolerance	3.0%	0.8%	0.6%
Precision (CV%)	0.21	0.11	0.09
Imprecision Tolerance	0.6%	0.3%	0.2%
Test Uncertainty (± µL / ± %)	0.03 / 0.15%	0.07 / 0.07%	0.11 / 0.06%

Test Result

PASS	PASS	PASS
------	------	------

PASS grade denotes that tolerances have been met according to OEM specifications. When accounting for measurement uncertainty, results may be outside of reported tolerance limits. The reported expanded uncertainty value uses a coverage factor k=2 to a coverage probability of approximately 95%.

Procedure Referenced: SOP02 REV13

V2.26