

# BRAND<sup>®</sup> CLASS A, USP, VOLUMETRIC GLASSWARE

Comprehensive line of Class A, USP, Certified volumetric glassware.

Precision analysis requires measuring instruments with consistently high precision. BLAUBRAND<sup>®</sup> volumetric instruments represent the highest degree of technical perfection. They are made of glass which is highly resistant to nearly all chemical substances - a prerequisite for durable precision. Every BLAUBRAND<sup>®</sup> volumetric instrument

is individually calibrated in BRAND's DIN EN ISO 9001 certified facility. Statistical process control ensures with a certainty of at least 99.6% (AQL≤0.4) compliance with predefined specifications.

## BLAUBRAND<sup>®</sup>—quality from the very start

- Accurate measurement instruments for precision analysis for companies audited by US authorities (such as the FDA)
- BLAUBRAND<sup>®</sup> USP volumetric instruments comply with the Class A error limits required by the United States Pharmacopeia
- USP batch certificate supplied, and USP individual certificate or DAkkS calibration available upon request
- Volumetric instruments are individually calibrated on computer controlled, automated production lines for maximum precision
- High quality inks, combined with carefully controlled annealing processes, provide durable graduations and long service life
- Thermal stress in the glass blanks is eliminated prior to adjustment so there are no permanent volume changes after heating, up to 250°C



**BRANDTECH<sup>®</sup>**  
SCIENTIFIC, INC.

Toll Free: 888-522-2726  
[www.brandtech.com](http://www.brandtech.com)

**BLAUBRAND® bulb pipette, 1 mark, class A, USP, Certified**

Volume (mL)	Error limit (+/-mL)	Qty./Pack	Cat. No.	2018 List Price
0.5	0.005	12	30601	\$103.30
1	0.006	12	30602	94.20
2	0.006	12	30603	105.50
3	0.01	6	30605	66.00
4	0.01	6	30606	85.65
5	0.01	6	30607	66.55
6	0.01	6	30608	91.40
7	0.01	6	30609	91.40
8	0.02	6	30610	91.40
9	0.02	6	30611	91.40
10	0.02	6	30612	67.50
15	0.03	6	30613	79.80
20	0.03	6	30614	86.25
25	0.03	6	30615	88.05
30	0.03	6	30616	100.90
40	0.05	6	30617	123.25
50	0.05	6	30618	112.40
100	0.08	6	30619	156.25

**BLAUBRAND® graduated pipette, class A, USP, Certified, cotton plug**

Volume (mL)	Subdivision (mL)	Error limit (+/-mL)	Qty./Pack	Cat. No.	2018 List Price
1	0.01	0.01	12	27506	\$77.60
2	0.02	0.01	12	27509	77.60
5	0.05	0.02	12	27511	95.05
10	0.1	0.03	12	27513	96.60

**BLAUBRAND® volumetric flask, class A, USP, Certified, Boro 3.3, glass stopper**

Volume (mL)	Error limit (+/-mL)	Stopper NS	Qty./Pack	Cat. No.	2018 List Price
5	0.02	10/19	2	36968	\$48.90
10	0.02	10/19	2	36973	48.90
20	0.02	10/19	2	36975	49.30
25	0.03	10/19	2	36977	49.30
50	0.05	12/21	2	36978	52.80
100	0.08	14/23	2	36979	57.20
200	0.1	14/23	2	36980	75.80
250	0.12	14/23	2	36981	79.00
500	0.2	19/26	2	36982	94.00
1000	0.3	24/29	1	36983	75.20
2000	0.5	29/32	1	36984	100.10

**BLAUBRAND® Amber volumetric flask, class A, USP, Certified, Boro 3.3, glass stopper**

Volume (mL)	Error limit (+/-mL)	Stopper NS	Qty./Pack	Cat. No.	2018 List Price
5	0.02	10/19	2	37461	\$103.80
10	0.02	10/19	2	37462	103.80
20	0.02	10/19	2	37463	103.60
25	0.03	10/19	2	37464	103.60
50	0.05	12/21	2	37465	107.60
100	0.08	14/23	2	37467	108.60
200	0.1	14/23	2	37468	134.70
250	0.12	14/23	2	37469	149.40
500	0.2	19/26	2	37470	181.70
1000	0.3	24/29	1	37471	131.20

**BLAUBRAND® graduated cylinder, tall form, class A, USP, Certified, Boro 3.3**

Volume (mL)	Subdivision (mL)	Error limit (+/-mL)	Height (mm)	Qty./Pack	Cat. No.	2018 List Price
5	0.1	0.05	115	2	32805	\$32.70
10	0.2	0.1	140	2	32808	32.70
25	0.5	0.25	170	2	32820	33.74
50	1	0.5	200	2	32828	38.23
100	1	0.5	260	2	32838	40.01
250	2	1	335	2	32848	76.14
500	5	2.5	365	2	32854	131.61
1000	10	5	465	1	32862	92.33
2000	20	10	505	1	32864	171.30

